

AMENDMENT OF SOLICITATION/MODIFICATION OF CONTRACT				1. CONTRACT ID CODE	PAGE OF PAGES	
				J	1	2
2. AMENDMENT/MODIFICATION NO. 0001		3. EFFECTIVE DATE 26-May-2015	4. REQUISITION/PURCHASE REQ. NO.		5. PROJECT NO. (If applicable)	
6. ISSUED BY CODE COMMANDING OFFICER - NAVFAC MARIANAS PSC 455, BOX 195 FPO AP GU 96540-2937		N40192	7. ADMINISTERED BY (If other than item 6) CODE See Item 6			
8. NAME AND ADDRESS OF CONTRACTOR (No., Street, County, State and Zip Code)				X	9A. AMENDMENT OF SOLICITATION NO. N40192-15-Q-4493	
				X	9B. DATED (SEE ITEM 11) 07-May-2015	
					10A. MOD. OF CONTRACT/ORDER NO.	
					10B. DATED (SEE ITEM 13)	
CODE		FACILITY CODE				
11. THIS ITEM ONLY APPLIES TO AMENDMENTS OF SOLICITATIONS						
<input checked="" type="checkbox"/> The above numbered solicitation is amended as set forth in Item 14. The hour and date specified for receipt of Offer <input type="checkbox"/> is extended, <input checked="" type="checkbox"/> is not extended. Offer must acknowledge receipt of this amendment prior to the hour and date specified in the solicitation or as amended by one of the following methods: (a) By completing Items 8 and 15, and returning _____ copies of the amendment; (b) By acknowledging receipt of this amendment on each copy of the offer submitted; or (c) By separate letter or telegram which includes a reference to the solicitation and amendment numbers. FAILURE OF YOUR ACKNOWLEDGMENT TO BE RECEIVED AT THE PLACE DESIGNATED FOR THE RECEIPT OF OFFERS PRIOR TO THE HOUR AND DATE SPECIFIED MAY RESULT IN REJECTION OF YOUR OFFER. If by virtue of this amendment you desire to change an offer already submitted, such change may be made by telegram or letter, provided each telegram or letter makes reference to the solicitation and this amendment, and is received prior to the opening hour and date specified.						
12. ACCOUNTING AND APPROPRIATION DATA (If required)						
13. THIS ITEM APPLIES ONLY TO MODIFICATIONS OF CONTRACTS/ORDERS. IT MODIFIES THE CONTRACT/ORDER NO. AS DESCRIBED IN ITEM 14.						
A. THIS CHANGE ORDER IS ISSUED PURSUANT TO: (Specify authority) THE CHANGES SET FORTH IN ITEM 14 ARE MADE IN THE CONTRACT ORDER NO. IN ITEM 10A.						
B. THE ABOVE NUMBERED CONTRACT/ORDER IS MODIFIED TO REFLECT THE ADMINISTRATIVE CHANGES (such as changes in paying office, appropriation date, etc.) SET FORTH IN ITEM 14, PURSUANT TO THE AUTHORITY OF FAR 43.103(B).						
C. THIS SUPPLEMENTAL AGREEMENT IS ENTERED INTO PURSUANT TO AUTHORITY OF:						
D. OTHER (Specify type of modification and authority)						
E. IMPORTANT: Contractor <input type="checkbox"/> is not, <input type="checkbox"/> is required to sign this document and return _____ copies to the issuing office.						
14. DESCRIPTION OF AMENDMENT/MODIFICATION (Organized by UCF section headings, including solicitation/contract subject matter where feasible.) Adding QUESTIONS/INQUIRIES AND ANSWERS.						
Except as provided herein, all terms and conditions of the document referenced in Item 9A or 10A, as heretofore changed, remains unchanged and in full force and effect.						
15A. NAME AND TITLE OF SIGNER (Type or print)			16A. NAME AND TITLE OF CONTRACTING OFFICER (Type or print)			
			TEL:		EMAIL:	
15B. CONTRACTOR/OFFEROR (Signature of person authorized to sign)		15C. DATE SIGNED	16B. UNITED STATES OF AMERICA BY _____ (Signature of Contracting Officer)		16C. DATE SIGNED 26-May-2015	

SECTION SF 30 BLOCK 14 CONTINUATION PAGE

SUMMARY OF CHANGES

(End of Summary of Changes)

N40192-15-Q-4493 (Questions/Inquiries and Answers):

1. The eductors with 1.5 fittings are available for 3 different flow rates -- 60 gpm, 90 gpm, or 120 gpm. Which do you prefer?

- The eductor flow rate should be adjustable

2. What does your end user intend to measure with the refractometer?

- The refractometer is for the foam.

3. I carry the Chemguard Class A foam instead of the Phos-Check Class A Foam. Am I allowed to bid the Chemguard Class A foam as a substitute to the Phos-Check Class A foam?

- I will need the specs for the chemguard foam to review, one of the purposes for the class a foam is suppression of wild land fires, so it needs to be environmentally friendly.

4. If you mod the solicitation, can you change the TFT description to say Eductor instead of Educator?

- The mod for spelling is ok educator.

5. I am familiar with DLA shipping requirements and the FDT program. Are they going to require MIL-STD- 129 labels and ISPM-15 pallets?

- As far as shipping requirements I am not very familiar with the requirements, requirement is the product is delivered to the end user without incurring any further cost.

6. I need to know what flow capacity you need on the TFT inline foam educator at; 65GPM, 95GPM, or 125GPM.

- 95GPM.

7.the following handheld refractometers which meet the test range of 0-10% with 1% increments so do you have a preference on which model as the PAL-1 is by far the most expensive and the MASTER-10M is the least expensive by far?

- \$130.00 is budgeted for this...."take which ever we can afford without going over budget."

- Unit should be an American Optical Model 10400 or 10441, Atago NI or equivalent.

8. What their sample is/what they are measuring?

- Firefighting foam, measure refractive index in accordance with NFPA 11.

9. What they are currently using:

- **No testing is being conducted.**

10. If they need an analog or a digital:

- **Digital.**

11. If they want temperature compensation or water resistance:

- **Water resistance.**

12. If they can accept a unit that reads more than 10%:

- **Yes.**